


# Certificate of Compliance

Product : Medical Grade Mobile Tablet PC

Applicant : MACTRON GROUP CO., LTD

Rating(s) : Adapter Rating:  
I/P: 100-250Vac, 50/60Hz, 1.5-0.75A max.  
O/P: 19Vdc, 3.15A  
Medical Grade Mobile Tablet PC rating: 19Vdc, 3.15A

Model/Type : MMS1016, MAS1016, MMSXXXX, MASXXXX (the first three X are for touch screen size: can be numeric 0 to 9, the last X for marketing used only: can be numeric 0 to 9)


Trademark : 

Test Report No. : M230320-05-3X-A0

Standard(s) : EN 60601-1:2006 + A11: 2011 + A1: 2013 + A12: 2014;

Applied : IEC 60601-1:2005, COR1:2006, COR2:2007, AMD1:2012  
(or IEC 60601-1:2012 reprint)

We hereby declare that the above equipment has been tested in our facility with the listed standard and found in compliance with the harmonized Standards of Medical Device Regulation 2017/745



Tony Hsu / Manager

2023-04-12



**Prodigy Technology Consultant Co.Ltd.**

No. 12, Gong 7th Rd., Linkou District, New Taipei City 24450, Taiwan Chinese Taipei  
TEL: 886-2-2603-7288/FAX: 886-2-2602-0908