




VERIFICATION OF COMPLIANCE

This is to certify that the product listed below was (were) tested in the BTL EMC Laboratory to comply with the required criteria levels of the follow-mentioned Generic Standards or Product Family Standard(s) and/or Basic Standard(s) based-on the essential conformity requirements of MDR Regulation (EU) 2017/745.

Equipment Medical Grade Mobile Tablet PC
Model Name MMS1016, MMSXXXX (the first three X are for touch screen size: can be numeric 0 to 9, the last X for marketing used only: can be numeric 0 to 9)
MAS1016, MASXXXX (the first three X are for touch screen size: can be numeric 0 to 9, the last X for marketing used only: can be numeric 0 to 9)
Brand Name 
Applicant MACTRON GROUP CO., LTD.
Address 2F, No. 86, Xinhua 1st Rd., Neihu Dist., Taipei City 114, Taiwan (R.O.C.)
Standard(s) EN 60601-1-2:2015
IEC 60601-1-2:2014
(for the HOME HEALTHCARE ENVIRONMENT)
EN IEC 61000-3-2:2019 Class D
EN 61000-3-3:2013+A1:2019
Report(s) BTL-EMC-1-2303T048

The test data, data evaluation, and equipment configuration contained in our test report(s) above was (were) obtained utilizing the test procedures, test instruments, test sites that has been accredited by the Authority of TAF according to the ISO/IEC 17025 quality assessment standard and technical standard(s). The test data contained in the referenced test report relate only to the EUT sample and item(s) tested.

Josh Lin

Authorized Signatory



BTL INC.

No.18, Ln. 171, Sec. 2, Jiuzong Rd.,
Neihu Dist., Taipei City 114, Taiwan
TEL:+886-2-2657-3299
FAX:+886-2-2657-3331

